

For Veterinary Use Only

READ ALL INSTRUCTIONS BEFORE BEGINNING THE TEST

RIDX™ CPV/CCV/Giardia Ag 3Dx Kit

[CAT No.: CGM-CTG-32, CGM-CTG-33]

Principles

The RIDX™ CPV/CCV/Giardia Ag 3Dx Kit is a lateral flow chromatographic immunoassay for the qualitative detection of CPV (Canine parvovirus), CPV (Canine coronavirus) and *Giardia* antigens in canine feces.

This kit shows two letters which are the test (T) line and the control (C) line for each test on the surface of the device. If the pathogenic antigens (CPV, CCV or *Giardia*) exist in the sample, that bind to the gold-conjugated pathogens (CPV, CCV or *Giardia*) specific antibodies. The antigen-antibody complex moves through the membrane by capillary force and responds to the pathogen (CPV, CCV or *Giardia*) specific antibodies on the test line, resulting in a red line. The control line indicates that the test is performed correctly and should appear when the test is complete.

The highly selective and sensitive monoclonal antibodies to CPV (CCV or *Giardia*) is used as capture and detector in the kit. The RIDX™ CPV/CCV/Giardia Ag 3Dx Kit can detect CPV antigens, CCV antigens and *Giardia* antigens in canine feces with high accuracy.

Performances

[CPV Ag Test]

1. Sensitivity & Specificity

		PCR		Total
		+	-	
RIDX™	+	101	0	101
CPV Ag	-	2	214	216
Test	Total	103	214	317

Sensitivity: 98.06% (101/103, 95% CI*: 93.19% ~ 99.47%)

Specificity: 100% (214/214, 95% CI: 98.24% ~ 100%)

Diagnostic Agreement: 99.37% (315/317, 95% CI: 97.73% ~ 99.83%)

* CI: Confidence Interval

2. Limit of Detection: 5×10^3 TCID₅₀/mL

3. CPV-2, CPV-2a, CPV-2b diagnosis are also available.

4. Cross-Reactivity

Potentially cross-reactive substances listed below have no effect on the performance of the RIDX™ CPV Ag Test.

Pathogen	Titer	Result
Canine coronavirus	1.00×10^6 TCID ₅₀ /mL	Negative
Canine distemper virus	1.00×10^5 TCID ₅₀ /mL	Negative
Canine influenza virus	1.00×10^6 EID ₅₀ /mL	Negative
<i>Escherichia coli</i>	3.56×10^8 CFU/mL	Negative
<i>Giardia</i> spp.	1.42×10^5 Cysts/μL	Negative
<i>Salmonella</i> spp.	1.00×10^6 CFU/mL	Negative

[CCV Ag Test]

1. Sensitivity & Specificity

		RT-PCR		Total
		+	-	
RIDX™	+	56	3	59
CCV Ag	-	3	118	121
Test	Total	59	121	180

Sensitivity: 94.92% (56/59, 95% CI*: 86.06% ~ 98.26%)

Specificity: 97.52% (118/121, 95% CI: 92.92% ~ 99.15%)

Diagnostic Agreement: 96.67% (174/180, 95% CI: 92.92% ~ 98.46%)

* CI: Confidence Interval

2. Limit of Detection: 1×10^5 TCID₅₀/mL

3. Cross-Reactivity

Potentially cross-reactive substances listed below have no effect on the performance of the RIDX™ CCV Ag Test Kit.

Pathogen	Titer	Result
Canine distemper virus	1.00×10^5 TCID ₅₀ /mL	Negative
Canine influenza virus	1.00×10^6 EID ₅₀ /mL	Negative
Canine parvovirus	1.00×10^6 TCID ₅₀ /mL	Negative
<i>Escherichia coli</i>	3.56×10^8 CFU/mL	Negative
<i>Giardia</i> spp.	1.42×10^5 Cysts/μL	Negative
<i>Salmonella</i> spp.	1.00×10^6 CFU/mL	Negative

[Giardia Ag Test]

1. Clinical Sensitivity & Clinical Specificity

		PCR		Total
		+	-	
RIDX™	+	34	4	38
Giardia Ag	-	1	109	110
Test	Total	35	113	148

Clinical Sensitivity: 97.14% (34/35, 95% CI*: 95.47% ~ 99.49%)

Clinical Specificity: 96.46% (109/113, 95% CI: 91.25% ~ 98.61%)

Diagnostic Accuracy: 96.62% (143/148, 95% CI: 92.34% ~ 98.55%)

* CI: Confidence Interval

2. Limit of Detection: 1.25 Cysts/μL

3. Cross-Reactivity

Potentially cross-reactive substances listed below have no effect on the performance of the RIDX™ Giardia Ag Test.

Pathogen	Titer	Result
Canine coronavirus	1.00×10^6 TCID ₅₀ /mL	Negative
Canine distemper virus	1.00×10^5 TCID ₅₀ /mL	Negative
Canine influenza virus	1.00×10^6 EID ₅₀ /mL	Negative
Feline calici virus	1.00×10^5 TCID ₅₀ /mL	Negative
Feline coronavirus	1.97×10^4 TCID ₅₀ /mL	Negative
Feline parvovirus	$1.00 \times 10^{5.5}$ TCID ₅₀ /mL	Negative

Kit Components


Component	Quantity/kit by CAT No.	
	CGM-CTG-32	CGM-CTG-33
1 CPV/CCV/ <i>Giardia</i> Ag 3Dx device	10	2
2 Dilution buffer for CPV & CCV tests	10	2
3 Dilution buffer for <i>Giardia</i> test	10	2
4 Disposable swab	20	4
5 Disposable dropper	20	4
6 Instructions for use	1	1

Storage & Stability

1. Store the test kit at 2~30°C (35.6~86.0°F). Do NOT freeze.
2. Do not store the test kit in direct sunlight.
3. The test kit is stable within the expiration date marked on the package label.


Sample Preparation

1. Canine fecal swab should be used for this test.
2. The samples should be tested immediately after collection.
3. If samples cannot be tested immediately, they should be stored at 2~8°C (35.6~46.4°F) for up to 24 hours. For longer storage, freeze at -20°C (-4°F) or below. Frozen samples should be brought to room temperature (15~30°C/59~86°F) before use.
4. The amount of fecal sample with swab may affect the results. It is required to follow the swab amount of feces as shown in the picture below. The excessive fecal amount may induce a false positive result and slow migration.



Test Procedure

1. All reagents and samples must be at room temperature (15~30°C /59~86°F) before use.
2. Collect feces samples using the swabs.
3. Put the first swab into the CCV/CPV Dilution Buffer tube and the second swab into the *Giardia* Dilution Buffer tube.
4. Stir the solution in each tube with the swab to disperse the sample into the buffer (approximately 10 seconds).
5. Remove the swab from each sample dilution buffer tube.
6. Wait for 20 seconds to settle down the large particles.
7. Remove a test device from the pouch and place it on a flat and dry surface.
8. Take the supernatant samples in each tube by using a disposable dropper.
9. Apply 4 drops of the mixed CPV/CCV sample solution into the sample holes (S1 and S2) for each, drop by drop vertically.
10. Apply 4 drops of the mixed *Giardia* sample solution into the sample hole (S3), drop by drop vertically.
11. Read test results at 10 minutes.




[Summary of Test Procedure]

Interpretation of Results


1. Positive Results

Test (T) line and control (C) line within the result window indicate the presence of pathogenic antigens.


[CPV positive]



[CCV positive]




[Giardia positive]



2. Negative Result

Only control (C) line appears in the result window.



3. Invalid Results

If the control (C) line does not appear, the result might be considered invalid. The sample should be retested.



Precautions

1. This test kit is for veterinary *in vitro* diagnostic use only for dogs. Do not use this test kit for other animals.
2. The test device is sensitive to humidity and heat. Use the test device within 10 minutes after removing the foil pouch.
3. Do not touch the membrane of the test device.
4. Do not use the test device if the foil pouch is damaged or the seal is open.
5. Do not use an expired test kit. The expiration date is marked on the package label.
6. Do not reuse the test components (device, buffer, dropper, and swab).
7. Do not mix components from different lot numbers because the components in this kit have been quality control tested as a standard batch unit.
8. Decontaminate and dispose of all samples, used kits, and potentially contaminated materials in the accordance with national and local regulations.
9. All samples should be handled as being potentially infectious. Wear protective gloves while handling samples. Wash hands thoroughly afterward.



SKYER, INC.

#532, 416, Hwagok-ro, Gangseo-gu, Seoul, 07548,
Republic of Korea
TEL: +82-2-706-6801, FAX: +82-50-4096-6988
Technical Support: marketing@skyer.co.kr
www.skyerdiagnostics.com

Korean Veterinary Diagnostics Manufacturer License No. 300